

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130172</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Freddie Wells</u>	ID# <u>4246</u> Date <u>11-14-11</u>
A Agency <u>Alaska State Troopers</u>	Phone # <u>907-883-5111</u>
Instrument Location <u>TOK, Alaska</u>	
Alco S/N <u>X173017</u> Target Value <u>.083</u> High Pressure <u>600</u>	
B Alco Test Values	
<u>.082</u> 1st Alco	<u>.084</u> 2nd Alco
Signature <u></u>	

(OVER)

SRMS  
11/22/11


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Feb, 2012 AK

 (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

NOVEMBER 14, 2011

OPERATOR'S NAME:  
WELLS/FREDDIE/LEE  
OPERATOR'S NUMBER: 4246  
SUBJECT'S LAST NAME:  
UOC

SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: 1230000  
DEPT/AGENCY: TOKO  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .083  
ALCO S/N: X173017

— BREATH ANALYSIS —

.083 ADJUSTED FOR 28.26 in  
ALCO TARGET .078 15:03  
BLANK TEST .000 15:05  
INTERNAL STANDARD VERIFIED 15:05  
ALCO TV 28.26 in .082 15:05  
BLANK TEST .000 15:06  
SUBJECT SAMPLE .000 15:06  
BLANK TEST .000 15:07  
ALCO TV 28.26 in .084 15:07  
BLANK TEST .000 15:08

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

NOVEMBER 14, 2011  
TIME 15:09

— DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 42c  
BAROMETER: 28.26 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢