

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130172</u> ✓
Name	<u>Brent Johnson</u>	ID# <u>4980</u> Date <u>3/6/08</u> ✓
A Agency	<u>AST</u>	Phone # <u>883-5111</u>
Instrument Location	<u>AST Tok</u>	
B Alco S/N	<u>Y301526</u> ✓	Target Value <u>.076</u> High Pressure <u>400</u>
Alco Test Value Average	<u>0.074</u> 1 <sup>st</sup> Alco <u>CRB</u>	<u>0.074</u> 2 <sup>nd</sup> Alco <u>CRB</u>
Signature	<u>[Signature]</u>	<u>CRB</u> <u>3/17/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brent Johnson, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature] 3/18/08  
Date  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

MARCH 06, 2000

OPERATOR'S NAME: <sup>CAS</sup> JOHNSON BRENT W  
OPERATOR'S NUMBER: 4900  
SUBJECT'S LAST NAME: CAL  
SUBJECT'S FIRST NAME/MI: VERIF OF  
O.L. #: 00000  
DEPT/AGENCY: TOKO  
CASE/REPORT: 00-000  
TEST TYPE: U  
ALCO TARGET VALUE: .076  
ALCO S/N: X301526

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.05 in  
ALCO TARGET .073 14:38  
BLANK TEST .000 14:39  
INTERNAL STANDARD VERIFIED 14:39  
ALCO TV 29.05 in .074 14:39  
BLANK TEST .000 14:40  
SUBJECT SAMPLE .000 <sup>CAS</sup> 14:40  
BLANK TEST .000 <sup>3/17/00</sup> 14:41  
ALCO TV 29.05 in .074 14:41  
BLANK TEST .000 14:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

MARCH 06, 2000  
TIME 14:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.05 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;{=}?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~