

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JUNE 11, 2008 *CS 6/15/08*

OPERATOR'S NAME:
JOHNSON BRENT W
OPERATOR'S NUMBER: 4900
SUBJECT'S LAST NAME:
CAL
SUBJECT'S FIRST NAME/MI :
VER OF
O.L. #: 000000
DEPT/AGENCY: TOK9
CASE/REPORT: 00-0000
TEST TYPE: V
ALCO TARGET VALUE: .076 ✓
ALCO S/N: X301526 ✓

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.62 in

ALCO TARGET	.075	16:52
BLANK TEST	.000	16:53
INTERNAL STANDARD	VERIFIED	16:53
ALCO TV 29.62 in	.075 ✓	16:54
BLANK TEST	.000	16:54
SUBJECT SAMPLE	.000 <i>CS 6/15/08</i>	16:55
BLANK TEST	.000	16:56
ALCO TV 29.62 in	.076 ✓	16:56
BLANK TEST	.000	16:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JUNE 11, 2008
TIME 16:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.62 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! "# \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢