

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130172</u> ✓	
Name	<u>Brent Johnson</u>	ID#	<u>4990</u> Date <u>8/7/08</u> ✓
A Agency	<u>AST</u>	Phone #	<u>883-5111</u>
Instrument Location	<u>Tok</u> ✓		
B Alco S/N	<u>V301526</u> ✓	Target Value	<u>.076</u> High Pressure <u>350</u>
Alco Test Value Average	<u>.074</u>	<u>.078</u>	
	1 st Alco	2 nd Alco	
Signature	<u>[Signature]</u>		<u>CS</u> <u>8/12/08</u>
(OVER)			

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brent Johnson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

8/13/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172V

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

AUGUST 07, 2000

OPERATOR'S NAME: JOHNSON BRENT W
OPERATOR'S NUMBER: 4980
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME/MI: OF CAL
O.L. #: 0000000
DEPT/AGENCY: TOKO
CASE/REPORT: 00-0000
TEST TYPE: U
ALCO TARGET VALUE: .076V
ALCO S/N: X301526V

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.51 in
ALCO TARGET .074 12:54
BLANK TEST .000 12:55
INTERNAL STANDARD VERIFIED 12:55
ALCO TU 29.51 in .074V 12:55
BLANK TEST .000 12:56
SUBJECT SAMPLE .000 12:57
BLANK TEST .000 12:57
ALCO TU 29.51 in .076 12:58
BLANK TEST .000 12:59

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

AUGUST 07, 2000
TIME 12:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~