

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

OCT 22 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

TOK

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130172 ✓

Name TIM Tuckwood ID# 3446 Date 10-9-09

A Agency Alaska state troopers Phone # 895-4801

Instrument Location TOK AST ✓

B Alco S/N X173017 Target Value .083 High Pressure 750

Alco Test Value Average .084 ✓ .087 ✓
 1st Alco 2nd Alco

Signature [Handwritten Signature] CBS
10/27/09

(OVER)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 6th day of NOV., 2009.
[Handwritten Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

OCTOBER 09, 2009

OPERATOR'S NAME: *CBS 10107109*
BUCKWOOD TIM
OPERATOR'S NUMBER: 3046
SUBJECT'S LAST NAME:
WERTZ
SUBJECT'S FIRST NAME (or I):
OF POL
SAL. #: 10045
DEPT/AGENCY: T010
CASE/REPORT #: 23767
TEST RESULT: ✓
ALCO. TARGET VALUE: 0.00 ✓
S/N: 130172

--- BREATH RESULTS ---

0.00 ADJUSTED FOR 29.91 in ✓
ALCO. TEST: 0.00 10:56
BLANK TEST: 0.00 10:57
INTERNAL STANDARD: VERIFIED 10:57
0.00 TO 29.91 in 0.04 ✓ 10:57
BLANK TEST: 0.00 ✓ 10:58
SUBJECT SAMPLE: *CBS 10107109* 0.00 ✓ 10:58
BLANK TEST: 0.00 ✓ 10:59
ALCO. TO 29.91 in 0.02 11:00
BLANK TEST: 0.00 11:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

OCTOBER 09, 2009 ✓
TIME 11:00

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
FIRMWARE DATE: 20/09/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
AIR SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO. TEST: 0.00 11:00
INTERNAL STANDARD: VERIFIED 11:00
0.00 TO 29.91 in 0.04 11:00
BLANK TEST: 0.00 11:01