VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

Supervisor/Operator Performing the Verification Procedure:

Name Brent Johnson ID# C1980 Date 10/14/08

A

Agency AST Phone # 883-5111

Instrument Location Tok

B

Alco S/N X173017 Target Value .083 High Pressure 1100

Alco Test Value Average .083 (1st Alco) .083 (2nd Alco)

Signature BJ

CB 11/4/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brent Johnson , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

11/5/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

OCT 20 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

OCTOBER 14, 2008 CDB 11/4/08

OPERATOR'S NAME:
JOHNSON BRENT W
OPERATOR'S NUMBER: 4980
SUBJECT'S LAST NAME:
CALIBRATION
SUBJECT'S FIRST NAME/MI :
VERIF OF
O.L. #: 000000
DEPT/AGENCY: TOK0
CASE/REPORT: 00-000
TEST TYPE: V
ALCO TARGET VALUE: .083
ALCO S/N: X173017

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.18 in

ALCO TARGET	.080	13:18
BLANK TEST	.000	13:19
INTERNAL STANDARD	VERIFIED	13:19
ALCO TV 29.18 in	.083	13:20
BLANK TEST	.000	13:20
SUBJECT SAMPLE	.000 CDB 11/4/08	13:21
BLANK TEST	.000	13:22
ALCO TV 29.18 in	.083	13:22
BLANK TEST	.000	13:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

OCTOBER 14, 2008
TIME 13:24

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.18 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~