

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130172</u> ✓	
Name <u>Brent W. Johnson</u>	ID# <u>4980</u> Date <u>12/1/08</u> ✓
Agency <u>AST</u>	Phone # <u>883-5111</u>
Instrument Location <u>Tok</u> ✓	
Alco S/N <u>X173017</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>1075</u>	
Alco Test Value Average <u>.084</u> ✓ <u>.086</u> ✓	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓	
(OVER) CSB 12/18/08	

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brent Johnson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/20/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

DECEMBER 07, 2008

OPERATOR'S NAME: *CSB 12/18/08*
JOHNSON BRENT W
OPERATOR'S NUMBER: 4980
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CAL
D.L. #: 00000
DEPT/AGENCY: TOKO
CASE/REPORT: 00-000
TEST TYPE: V ✓
ALCO TARGET VALUE: .083 ✓
ALCO S/N: X173017

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.57 in
ALCO TARGET .082 09:45
BLANK TEST .000 09:46
INTERNAL STANDARD VERIFIED 09:46
ALCO TV 29.57 in .084 ✓ 09:46
BLANK TEST .000 09:47
SUBJECT SAMPLE *CSB 12/18/08* .000 09:48
BLANK TEST .000 09:49
ALCO TV 29.57 in .086 ✓ 09:49
BLANK TEST .000 09:50

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

DECEMBER 07, 2008 ✓
TIME 09:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■