

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130173 ✓			
Name	<u>[Signature]</u>	ID#	3823 ✓			
		Date	3-24-11 ✓			
A	Agency	AST	Phone # 822-3263			
	Instrument Location	Glennallen				
B	Alco S/N	78820 ✓	Target Value	0.083 ✓	High Pressure	800
	Alco Test Values	0.080 ✓	0.080 ✓			
		1st Alco	2nd Alco			
	Signature	<u>[Signature]</u>				

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 1st day of APRIL, 2011.



[Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

MARCH 24, 2011

OPERATOR'S NAME:

HEINBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/NI :

UOC

O.L. #: UOC

DEPT/AGENCY: ENKD

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: 0.083

ALCO S/N: 78828

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.24 in

ALCO TARGET	.078	07:47
BLANK TEST	.000	07:48
INTERNAL STANDARD	VERIFIED	07:48
ALCO TV 28.24 in	.000	07:48
BLANK TEST	.000	07:49
SUBJECT SAMPLE	.000	07:50
BLANK TEST	.000	07:50
ALCO TV 28.24 in	.000	07:51
BLANK TEST	.000	07:52

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

MARCH 24, 2011

TIME 07:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 28.24 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢