

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 03 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130173

A  
Name J Heinbaugh ID# 5828 Date 2-10-09  
Agency AST Phone # 822-3263  
Instrument Location Glennallen

B  
Alco S/N 75602 Target Value 0.080 High Pressure SSO  
Alco Test Values 0.080 0.080  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco  
Signature [Signature] [Signature]  
8/6/09

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 14<sup>th</sup> day of August, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

### Breath Test Record

Serial Number 130173 Model DataMaster  
Location GKNO AST/Glennallen

*CS  
8/10/09*

### Subject Information

Last Name Voc O.L. # VOC  
First Name Voc

### Operator Information

Operator Name Heinbaugh Case VOC  
Operator # 5828 Department GKNO

### Test Results

Alco Target 0.080 Alco Serial 75602  
Nominal 0.076 Barometer 28.47 Test Type V  
Dry Gas 1 0.080 Time: 12:48:42 Test Date 02/10/2009  
Subject Alcohol 0.000 Time: 12:50:08 Test Time 12:47:07  
Dry Gas 2 0.080 Time: 12:51:12 Report Y

## DIAGNOSTIC CHECK

### Diagnostic Test Record

Serial Number 130173 Model DataMaster  
Location GKNO AST/Glennallen

### Diagnostic Test Results

Test Date 02/10/2009 Computer Okay  
Test Time 12:52:15 Program Okay  
Software Date 02/20/2001  
Sample Chamber (C) 49  
Breath Tube (C) 42  
Barometer (in) 28.47  
Flow Detector Okay  
High Speed Pump Okay  
Detector Okay  
Filter Okay  
Quartz Okay  
Calibration Okay