

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130173</u> ✓
Name		<u>Josh Heinbaugh</u>	ID# <u>5828</u> Date <u>3-25-09</u> ✓
A	Agency	<u>AST</u>	Phone # <u>822-3263</u>
Instrument Location		<u>Glennallen</u> ✓	
Alco S/N		<u>130173</u>	Target Value <u>0.080</u> ✓
		<u>75602</u> (3.2409 cup)	High Pressure <u>400</u>
B	Alco Test Values	<u>0.080</u> ✓	<u>0.080</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<u>[Signature]</u> ✓	<u>[Signature]</u> 3/24/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Josh Heinbaugh, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/28/09

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

MARCH 25, 2009 ✓  
OPERATOR'S NAME: WEINBAUGH  
OPERATOR'S NUMBER: 5828  
SUBJECT'S LAST NAME: VOC  
SUBJECT'S FIRST NAME/NI: VOC  
O.I.L. #: VOC  
DEPT/AGENCY: GSK9  
CASE/REPORT: VOC  
TEST TYPE: U  
ALCO TARGET VALUE: 0.000 ✓  
ALCO S/N: 75602 ✓

--- BREATH ANALYSIS ---

.050 ADJUSTED FOR 28.27 in		
ALCO TARGET	.075	15:05
BLANK TEST	.000	15:06
INTERNAL STANDARD	VERIFIED	15:06
ALCO TV 28.27 in	.000 ✓	15:06
BLANK TEST	.000	15:07
SUBJECT SAMPLE	.000	15:08
BLANK TEST	.000	15:09
ALCO TV 28.27 in	.000 ✓	15:09
BLANK TEST	.000	15:10

COB 3/27/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

MARCH 25, 2009 ✓  
TIME 15:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHANGER: 49c  
BREATH TUBE: 41c  
BAROMETER: 28.27 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmnop  
qrstuvwxy z{|}~