

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


| | | | |
|--|---------------------|---|--|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | 130173 ✓ |
| Name | José Heimbach | ID# | 5828 (31519) ✓ JDK8 |
| | | Date | 3-4-10 ✓ |
| A | Agency | AST | Phone # 800-3263 |
| | Instrument Location | Glennallen ✓ | |
| B | Alco S/N | X124664 ✓ | Target Value 0.083 ✓ |
| | | | High Pressure 1200 |
| | Alco Test Values | 0.079 ✓ 1 st Alco | 0.080 ✓ 2 nd Alco |
| | Signature |  ✓ |  3/11/10 |

(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of March, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173 ✓

MARCH 04, 2010 ✓ *COB 3/11/10*

OPERATOR'S NAME: HETENBAUGH
OPERATOR'S NUMBER: 5828
SUBJECT'S LAST NAME: UOC
SUBJECT'S FIRST NAME/MI : UOC
O.L. #: UOC
DEPT/AGENCY: GAKO
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: 0.083 ✓
ALCO S/N: X124669 ✓

— BREATH ANALYSIS —

.083 ADJUSTED FOR 27.88 in
ALCO TARGET .077 09:25
BLANK TEST .000 09:26
INTERNAL STANDARD VERIFIED 09:26
ALCO TV 27.88 in .079 ✓ 09:26
BLANK TEST .000 09:27
SUBJECT SAMPLE *COB 3/11/10* .000 09:28
BLANK TEST .000 09:29
ALCO TV 27.90 in ✓ .000 09:29
BLANK TEST .000 09:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173 ✓

MARCH 04, 2010 ✓
TIME 09:23 ✓

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 27.88 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■