

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

JUN 29 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130173</u> ✓
Name		<u>Heinbaugh</u>	ID# <u>5808</u> ✓ Date <u>6-27-10</u> ✓
A	Agency	<u>Glennallen AST</u>	Phone # <u>800-3263</u>
Instrument Location		<u>Glennallen AST</u>	
Alco S/N		<u>X124669</u> ✓	Target Value <u>0.083</u> ✓ High Pressure <u>700</u>
B	Alco Test Values	<u>0.081</u> ✓ 1 <sup>st</sup> Alco	<u>0.080</u> ✓ 2 <sup>nd</sup> Alco
Signature		<u>[Signature]</u>	<u>[Signature]</u> 7/8/10
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of July, 2010.

[Signature]  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATA MASTER cdm 130173

JUNE 27, 2010

OPERATOR'S NAME: HEINBAUGH  
OPERATOR'S NUMBER: 5828  
SUBJECT'S LAST NAME: UOC  
SUBJECT'S FIRST NAME/MI: UOC  
O.L. #: UOC  
DEPT/AGENCY: GSK0  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: 0.083  
ALCO S/N: X124669

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.34 in		
ALCO TARGET	.078	12:26
BLANK TEST	.000	12:27
INTERNAL STANDARD	VERIFIED	12:28
ALCO TV 28.34 in	.081	12:28
BLANK TEST	.000	12:29
SUBJECT SAMPLE	.000	12:29
BLANK TEST	.000	12:30
ALCO TV 28.35 in	.080	12:30
BLANK TEST	.000	12:31

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATA MASTER cdm 130173

JUNE 27, 2010  
TIME 12:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01

HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c

BAROMETER: 29.35 in

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~■