

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130173 ^{BMB} 12/8/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173

OCTOBER 13, 2010

OPERATOR'S NAME:

HEINBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: VOC

DEPT/AGENCY: ANKO

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: 0.083

ALCO S/N: X124669

— BREATH ANALYSIS —

.083 ADJUSTED FOR 28.34 in
ALCO TARGET .078 11:43
BLANK TEST .000 11:45
INTERNAL STANDARD VERIFIED 11:45
ALCO TV 28.34 in .081 11:45
BLANK TEST .000 11:46
SUBJECT SAMPLE .000 11:46
BLANK TEST .000 11:47
ALCO TV 28.34 in .081 11:48
BLANK TEST .000 11:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130173

OCTOBER 13, 2010

TIME 11:48

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 28.34 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()-./0123456789:;<=>?@#CDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~