

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130174</u>
Name	<u>Lowe, Gary</u>	ID# <u>7219</u> Date <u>4/15/2011</u>
A Agency	<u>Haines Police</u>	Phone # <u>766-2121</u>
Instrument Location	<u>Haines PD</u>	
Alco S/N <u>X301526</u>		Target Value <u>.078</u> High Pressure <u>1000</u>
B Alco Test Values	<u>.081</u> 1 <sup>st</sup> Alco	<u>.080</u> 2 <sup>nd</sup> Alco
Signature	<u>Gary R Lowe</u>	
(OVER)		


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011. NJB

Daniel L. Nicholas (Notary Seal Stamp)  
Daniel L. Nicholas KYLE KLOPPER  
Notary Public, State of Alaska  
Commission Expires with Office



CRF  
4/21/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

APRIL 15, 2011 *CRF 4/21/11*

OPERATOR'S NAME:  
LOWE, GARY  
OPERATOR'S NUMBER: 7219  
SUBJECT'S LAST NAME:  
TEST  
SUBJECT'S FIRST NAME/NI :  
VERI  
O.L. #: 04152011  
DEPT/AGENCY: HNS1  
CASE/REPORT: 11-415  
TEST TYPE: V  
ALCO TARGET VALUE: .078  
ALCO S/N: X301526

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.32 in		
ALCO TARGET	.079	00:54
BLANK TEST	.000	00:55
INTERNAL STANDARD	VERIFIED	00:55
ALCO TV 30.32 in	.001	00:55
BLANK TEST	.000	00:56
SUBJECT SAMPLE	.000	00:56
BLANK TEST	.000	00:57
ALCO TV 30.33 in	.000	00:57
BLANK TEST	.000	00:58

*CRF 4/21/11*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

APRIL 15, 2011  
TIME 08:59

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.33 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`\*abcdefghijklmnop  
qrstuvwxyz{|}~■