

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130174</u>
Name	<u>Gary Lowe</u>	ID# <u>7219</u> Date <u>11/15/11</u>
A Agency	<u>Haines P.D.</u>	Phone # <u>766-2121</u>
Instrument Location <u>Haines P.D.</u>		
B Alco S/N <u>4J</u> Target Value <u>.080</u> High Pressure _____		
Alco Test Values		
	<u>.075</u> 1 <sup>st</sup> Alco	<u>.076</u> 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>		

(OVER)

BMB  
11/22/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1<sup>st</sup> day of Feb, 2012 AK

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

NOVEMBER 15, 2011 *BMB 11/22/11*

OPERATOR'S NAME: LOWE, GARY  
OPERATOR'S NUMBER: 7219  
SUBJECT'S LAST NAME: TEST  
SUBJECT'S FIRST NAME/MI: VERIFY  
O.L. #: 12345  
DEPT/AGENCY: HNSI  
CASE/REPORT: 111511  
TEST TYPE: V  
ALCO TARGET VALUE: .000  
ALCO S/N: 4J

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.06 in		
ALCO TARGET	.000	09:25
BLANK TEST	.000	09:26
INTERNAL STANDARD	VERIFIED	09:26
ALCO TV 30.06 in	.075	09:26
BLANK TEST	.000	09:27
SUBJECT SAMPLE <i>BMB 11/22/11</i>	.000	09:27
BLANK TEST	.000	09:28
ALCO TV 30.06 in	.076	09:28
BLANK TEST	.000	09:29

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

NOVEMBER 15, 2011  
TIME 09:28

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
"##"()+-,./0123456789:;<=>@ABCDEFGHIJ  
KLMNOPQRSTU VWXYZ\]^\_`abcde fghijklmno  
pqrstuvwxy z{|}~|