VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

Supervisor/Operator Performing the Verification Procedure:

Name J Swartz ID# 4314 Date 1/16/08

A

Agency SCDL Phone # 269-5592

Instrument Location UAA PD

Alco S/N X124656 Target Value .075 High Pressure 400 psi

B

Alco Test Values .075 (1st Alco) .074 (2nd Alco)

Signature [signature] CB 4/2/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed, [signature] 4/2/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Breath Test Record

Serial Number: 130174 Model: DataMaster
Location: ANC7 University Police - UAA

CGB 4/2/08

Subject Information

Last Name: Verification Of Cal O.L. #: L
First Name: L

Operator Information

Operator Name: Swartz/J Case:
Operator #: 4314 Department: AAA1

Test Results

Alco Target	.075	Alco Serial	X124656		
Nominal	0.075	Barometer	30.24	Test Type	V
Dry Gas 1	0.075	Time:	16:07:53	Test Date	01/16/2008
Subject Alcohol	0.000	Time:	16:09:12	Test Time	16:06:38
Dry Gas 2	0.074	Time:	16:10:15	Report	Y

CGB 4/2/08

DIAGNOSTIC CHECK

Diagnostic Test Record

Serial Number: 130174 Model: DataMaster
Location: ANC7 University Police - UAA

Diagnostic Test Results

Test Date	01/16/2008	Computer	Okay
Test Time	16:05:06	Program	Okay
		Software Date	02/20/2001
		Sample Chamber (C)	49
		Breath Tube (C)	41
		Barometer (in)	30.24
		Flow Detector	Okay
		High Speed Pump	Okay
		Detector	Okay
		Filter	Okay
		Quartz	Okay
		Calibration	Okay