

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>J Swartz</u>	DataMaster cdm S/N <u>130174</u> ✓ ID# <u>4314</u> Date <u>03/07/09</u> ✓
A Agency <u>SCDL</u>	Phone # <u>269-5592</u>
Instrument Location <u>SCDL</u> ✓	
B Alco S/N <u>SC0708</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>900 psi</u>	
Alco Test Values <u>.080</u> ✓ <u>.080</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓	
(OVER)	

BMB
9/9/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number	130174	Model	DataMaster
Location	ANC7	University Police - UAA	

Subject Information

Last Name	Verification Of Cal	O.L. #	L
First Name	L		

Operator Information

Operator Name	Swartz/J	Case	
Operator #	4314	Department	AAA1

Test Results

Alco Target	0.080	Alco Serial	SCOTT08		
Nominal	0.081	Barometer	30.63	Test Type	V
Dry Gas 1	0.080	Time	20:45:37	Test Date	03/07/2009
Subject Alcohol	0.000	Time	20:46:43	Test Time	20:44:22
Dry Gas 2	0.080	Time	20:47:42	Report	Y

9/19/09
9/19/09

Diagnostic Test Record

Serial Number	130174	Model	DataMaster
Location	ANC7	University Police - UAA	

Diagnostic Test Results

Test Date	03/07/2009	Computer	Okay
Test Time	20:42:47	Program	Okay
Software Date	01/08/2009	Sample Chamber (C)	49
Breath Tube (C)	40	Barometer (in)	30.63
Flow Detector	Okay	High Speed Pump	Okay
Detector	Okay	Filter	Okay
Quartz	Okay	Calibration	Okay