

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130174

Name Kevin Stevenson ID# 2414 Date 05-23-08 ✓

A Agency UAA Police Department Phone # 786-1120

Instrument Location UAA Police Department ✓

B Alco S/N X301591 ✓ Target Value .080 ✓ High Pressure 100

Alco Test Value Average .079 ✓  
1<sup>st</sup> Alco .079 ✓  
2<sup>nd</sup> Alco

Signature [Signature] (Stevenson) CS 6/18/08

(OVER)

Under the Alaska Rules of Evidence, I certify that,

(Do Not write in the area below.)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Stevenson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

6/19/08

Date

MAY 29 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

MAY 23, 2008 ✓ *COB 6/19/08*

OPERATOR'S NAME:

STEVENSON KEVIN L

OPERATOR'S NUMBER: 2414

SUBJECT'S LAST NAME:

VERIFICATION:

SUBJECT'S FIRST NAME/NI:

VERIFICATION:

O.L. #: 1234

DEPT/AGENCY: ANC7

CASE/REPORT: 0009999

TEST TYPE: U

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X301591 ✓

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.37 in

ALCO TARGET	.001	00:49
BLANK TEST	.000	00:50
INTERNAL STANDARD	VERIFIED	00:50
ALCO TU 30.37 in	.079 ✓	00:51
BLANK TEST	.000	00:51
SUBJECT SAMPLE	.000 <i>COB 6/19/08</i>	00:52
BLANK TEST	.000	00:53
ALCO TU 30.37 in	.079 ✓	00:53
BLANK TEST	.000	00:54

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

MAY 23, 2008 ✓

TIME 00:56

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.37 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~