

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130174 ✓</u>	
Name <u>J Swartz</u>	ID# <u>4314</u> Date <u>7/18/08 ✓</u>
Agency <u>SCDL</u>	Phone # <u>269-5592</u>
Instrument Location <u>SCDL ✓</u>	
Alco S/N <u>SCOTT07 ✓</u> Target Value <u>.080 ✓</u> High Pressure <u>400 psi</u>	
Alco Test Values	<u>.078 ✓</u> <u>.078 ✓</u>
1 st Alco	2 nd Alco
Signature <u>J Swartz</u>	<u>CS</u> <u>7/28/08</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



J Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/30/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JULY 18, 2008 ✓
COB 7/18/08

OPERATOR'S NAME: SMARTZ/J
OPERATOR'S NUMBER: 4314
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:
L
O.L. #: L
DEPT/AGENCY: AAA1
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: SCOTT02 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.67 in

ALCO TARGET	.079	14:40
BLANK TEST	.000	14:41
INTERNAL STANDARD	VERIFIED	14:41
ALCO TU 29.67 in	.078 ✓	14:41
BLANK TEST	.000	14:42
SUBJECT SAMPLE	.000 ✓	14:43
BLANK TEST	.000 ✓	14:44
ALCO TU 29.67 in	.078 ✓	14:44
BLANK TEST	.000	14:45

8 7/18/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JULY 18, 2008 ✓
TIME 14:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c

BAROMETER: 29.67 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~#