

# VERIFICATION OF CALIBRATION REPORT

SEP 24 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

SEP 24 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130174 ✓

Name Gary Lowe ID# 7219 Date 09/21/2009 ✓

A Agency Haines PD Phone # 766-2063

Instrument Location Haines PD ✓

---

B Alco S/N X124631 ✓ Target Value .082 ✓ High Pressure 850

Alco Test Values .082 ✓ .082 ✓  
 1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Gary Lowe ✓

CDS  
9/28/09

(OVER)


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 28<sup>th</sup> day of Oct., 2009.

Carolyn M. Noland  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

SEPTEMBER 21, 2009 ✓

OPERATOR'S NAME:  
LOWE, GARY

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

GARY

SUBJECT'S FIRST NAME/MI :  
CAL

D.L. #: 092109

DEPT/AGENCY: HNS1

CASE/REPORT: 09212009

TEST TYPE: U

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X124631 ✓

### BREATH ANALYSIS

BAR ADJUSTED FOR 30.24 in		
ALCO TARGET	.002	11:00
BLANK TEST	.000	11:00
INTERNAL STANDARD	VERIFIED	11:09
ALCO TV 30.24 in	.002 ✓	11:09
BLANK TEST	.000	11:09
SUBJECT SAMPLE <i>COS</i>	.000	11:09
BLANK TEST <i>7/21/09</i>	.000	11:21
ALCO TV 30.24 in	.002 ✓	11:21
BLANK TEST	.000	11:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

SEPTEMBER 21, 2009 ✓

TIME 11:14

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/00/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.24 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKLMN  
OPQRSTUVWXYZ[\]^\_`abcdefg|hijklmnop  
qrstuvwxyz{|}~@