

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 25 2010

DataMaster cdm S/N <u>130174</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Gary Lowe</u>	ID# <u>7219</u> Date <u>01/19/2010</u>
A Agency <u>Haines Police</u>	Phone # <u>766-2121</u>
Instrument Location <u>Haines P.D.</u>	
Alco S/N <u>X124631</u> Target Value <u>.082</u> High Pressure <u>600 psi</u>	
B Alco Test Values <u>.080</u> <u>.080</u>	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25th day of January 2010.
[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174 ✓

JANUARY 19, 2010 ✓ *COB 1/25/10*

OPERATOR'S NAME:
LOWE, GARY, A

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:
TEST

SUBJECT'S FIRST NAME/MI :
ALCO

O.L. #: 01192010

DEPT/AGENCY: HNS1

CASE/REPORT: 2010-0119

TEST TYPE: U ✓

ALCO TARGET VALUE: .082 ✓

ALCO S/N: X124631 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.44 in		
ALCO TARGET	.000	12:17
BLANK TEST	.000	12:18
INTERNAL STANDARD	VERIFIED	12:18
ALCO TV 29.44 in	.000	12:18
BLANK TEST	.000	12:19
SUBJECT SAMPLE	.000	12:19
BLANK TEST	.000	12:20
ALCO TV 29.44 in	.000	12:20
BLANK TEST	.000	12:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174 ✓

JANUARY 19, 2010 ✓
TIME 12:14 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.44 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~