

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Lowe, Gary ID# 7219 ✓ Date 9/20/10 ✓

A Agency Haines Police Phone # 766-2121

Instrument Location Haines Police Station

B Alco S/N X301526 ✓ Target Value .078 ✓ High Pressure 1400

Alco Test Values .081 ✓ .080 ✓
 1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
9/27/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 28th day of Oct, 2010.

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

SEPTEMBER 28, 2010 *8mb*

OPERATOR'S NAME: *9/27/10*

LOWE, GARY

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

UER

O.L. #: 09282010

DEPT/AGENCY: HNSI

CASE/REPORT: 10-0920

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301526

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.22 in		
ALCO TARGET	.078	13:46
BLANK TEST	.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
ALCO TV 30.22 in	.081	13:48
BLANK TEST	.000	13:48
SUBJECT SAMPLE <i>8mb</i>	.000	13:49
BLANK TEST <i>9/27/10</i>	.000	13:49
ALCO TV 30.22 in	.000	13:50
BLANK TEST	.000	13:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

SEPTEMBER 28, 2010

TIME 13:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.22 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~||