

VERIFICATION OF CALIBRATION REPORT

NOV 30 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130174 ✓

Name Lowe, Gary ID# 7219 ✓ Date 11/19/2010 ✓

A Agency Haines Police Phone # 766-2121

Instrument Location Haines Police

B Alco S/N X301526 ✓ Target Value .078 ✓ High Pressure 1300

Alco Test Values .082 .081 ✓

~~.080~~ DLW 12/9/10 1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB 12/16/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of DEC, 2010.

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174
NOVEMBER 19, 2010
OPERATOR'S NAME: LOWE/GARY
OPERATOR'S NUMBER: 7219
SUBJECT'S LAST NAME: TEST
SUBJECT'S FIRST NAME/NI: VERY
O.L. #: 111910
DEPT/AGENCY: HNS1
CASE/REPORT: NOV192010
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: K301526

BMB 12/16/10

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.77 in		
ALCO TARGET	.080	14:05
BLANK TEST	.000	14:06
INTERNAL STANDARD	VERIFIED	14:06
ALCO TU 30.77 in	.082	14:06
BLANK TEST	.000	14:07
SUBJECT SAMPLE	.000	14:08
BLANK TEST	.000	14:08
ALCO TU 30.77 in	.081	14:09
BLANK TEST	.000	14:09

BMB 12/16/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174
NOVEMBER 19, 2010
TIME 14:10

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/00/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.77 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~