

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130175</u>
Name	<u>John Rhyshek</u>	ID#	<u>5670</u> Date <u>4-4-08</u>
A	Agency	<u>Houston Police Dept.</u>	Phone # <u>892-6447</u>
	Instrument Location	<u>Houston Police Dept.</u>	
B	Alco S/N	<u>X172991</u>	Target Value <u>.077</u> High Pressure <u>750 psi</u>
	Alco Test Values	<u>.073</u> <u>.077 CBS</u> 1 st Alco	<u>.072</u> <u>.077 CBS</u> 2 nd Alco
	Signature		<u>CBS</u> <u>4/5/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Rhyshek, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/20/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130175

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130175

APRIL 04, 2008

OPERATOR'S NAME: RHYSEK JOHN C
OPERATOR'S NUMBER: 5670
SUBJECT'S LAST NAME: RHYSEK
SUBJECT'S FIRST NAME/MI: JOHN C

D.L. #: 1013203
DEPT/AGENCY: HPD1
CASE/REPORT: 00-0000
TEST TYPE: U
ALCO TARGET VALUE: .077
ALCO S/N: X172991

BREATH ANALYSIS

ALCO ADJUSTED FOR 29.68 in
ALCO TARGET .076
BLANK TEST .000
INTERNAL STANDARD VERIFIED
ALCO TU 29.67 in .073
BLANK TEST .000
SUBJECT SAMPLE .000
BLANK TEST .000
ALCO TU 29.68 in .072
BLANK TEST .000

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130175

APRIL 04, 2008
TIME 16:37

DIAGNOSTIC CHECK

COUNTER: OKAY
FUNCTION: OKAY
EXPIRE DATE: 02/20/01
VALVE CHAMBER: 42c
BREATH TUBE: 42c
METER: 29.68 in
FLAME DETECTOR: OKAY
SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@