

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>1301875</u>
Name		<u>Colleen O'Bryant</u>	ID# <u>6/6337</u> Date <u>10/24/08</u>
A	Agency	<u>SCD</u>	Phone # <u>269-5723</u>
Instrument Location		<u>Houston RD</u>	
Alco S/N		<u>X172991</u>	Target Value <u>0.077</u> High Pressure <u>700psi</u>
B	Alco Test Values	<u>0.073</u> 1 <sup>st</sup> Alco	<u>0.073</u> 2 <sup>nd</sup> Alco
Signature		<u>C. O'Bryant</u>	<u>COB</u> <u>11/4/08</u>


(OVER)

Under the Alaska Rules of Evidence, I certify that (print name in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Colleen S O'Bryant, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

11/5/08

Date

OCT 27 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130175

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130175

OCTOBER 24, 2008 *2008/10/28*

OPERATOR'S NAME:  
O'BRYAN/COLLEEN/S  
OPERATOR'S NUMBER: 6237  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI:  
VOC  
O.I.L. #: 0123456  
DEPT/AGENCY: AAA1  
CASE/REPORT: 08-1234  
TEST TYPE: U  
ALCO TARGET VALUE: .077 ✓  
ALCO S/N: X172991 ✓

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.83 in		
ALCO TARGET	.076	17:32
BLANK TEST	.000	17:33
INTERNAL STANDARD	VERIFIED	17:33
ALCO TU 29.83 in	.073	17:34
BLANK TEST	.000 ✓	17:34
SUBJECT SAMPLE	.000 ✓	17:35
BLANK TEST	.000 ✓	17:36
ALCO TU 29.85 in	.073	17:36
BLANK TEST	.000	17:37

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130175

OCTOBER 24, 2008 ✓  
TIME 17:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.83 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
! "# \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~