

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130176 ✓

Name Patrick Smith ID# 6397 Date 28 Jan 09 ✓

A Agency Clear AFS Security Forces Phone # 585-6293

Instrument Location Bldg 37 Clear Air Force Station ✓


B Alco S/N X124609 ✓ Target Value .081 ✓ High Pressure 1250

Alco Test Values .084 ✓ .084 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature] ✓ CSB  
2/9/09

Under the Alaska Rules of Evidence, I certify that, <sup>(OVER)</sup>

- (a) I am <sup>(Do Not write in the area below)</sup> Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick J Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature] 2/9/09  
Date  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

JAN 30 2009

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

JANUARY 28, 2009 ✓  
*08/21/09*

OPERATOR'S NAME:  
SMITH/PATRICK/J  
OPERATOR'S NUMBER: 6397  
SUBJECT'S LAST NAME:  
UOC

SUBJECT'S FIRST NAME/MI :  
UOC

O.I.L. #: 1234567 ✓  
DEPT/AGENCY: CLR4  
CASE/REPORT: 20090128  
TEST TYPE: U  
ALCO TARGET VALUE: .081 ✓  
ALCO S/N: X124609

### — BREATH ANALYSIS —

.081 ADJUSTED FOR 29.65 in		
ALCO TARGET	.080	08:17
BLANK TEST	.000	08:18
INTERNAL STANDARD	VERIFIED	08:18
ALCO TU 29.65 in	.084 ✓	08:19
BLANK TEST	.000	08:20
SUBJECT SAMPLE	.080	08:20
BLANK TEST	.000	08:21
ALCO TU 29.65 in	.084 ✓	08:21
BLANK TEST	.000	08:22

*08/21/09*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

JANUARY 28, 2009 ✓  
TIME 08:14

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.65 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrsstuvwxyz{|}~#