DEC 0 8 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

Supervisor/Operator Performing the Verification Procedure:

A

Name CHRIS WILLIAMS ID# 7086 Date 24 NOV 08

Agency CLEAR SECURITY FORCES Phone # (907) 585-6313

Instrument Location BLDG 37 CLEAR AIR FORCE STATION, ALASKA

B

Alco S/N X124609 Target Value .081 High Pressure 1300 PSI

Alco Test Values .083 (1st Alco) .083 (2nd Alco)

Signature [signature] CCB 12/18/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris L Williams, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 12/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

NOVEMBER 24, 2008 CDB 12/18/08

OPERATOR'S NAME:
WILLIAMS/CHRIS/L
OPERATOR'S NUMBER: 7096
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
D.L. #: 1234567
DEPT/AGENCY: CLR4
CASE/REPORT: 757
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X124609

— BREATH ANALYSIS —

.081 ADJUSTED FOR 28.58 in

ALCO TARGET	.077	16:07
BLANK TEST	.000	16:08
INTERNAL STANDARD	VERIFIED	16:08
ALCO TV 28.58 in	.083	16:09
BLANK TEST	.000	16:10
SUBJECT SAMPLE	.000	16:10
BLANK TEST	.000	16:11
ALCO TV 28.58 in	.083	16:11
BLANK TEST	.000	16:12

CDB 12/18/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

NOVEMBER 24, 2008
TIME 15:58

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	28.56 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~