

# VERIFICATION OF CALIBRATION REPORT MAY 13 2011

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130176</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>DARYL REECE</u>	ID# <u>2088</u> Date <u>05/11/11</u>	
A Agency <u>SITKA POLICE DEPT.</u>	Phone # <u>747-3245</u>	
Instrument Location <u>304 LAKE STREET</u>		
B Alco S/N <u>X301572</u> Target Value <u>.078</u> High Pressure <u>250</u>		
Alco Test Values	<u>.080</u> 1st Alco	<u>.085</u> 2nd Alco
Signature <u>Daryl Rice</u>		
(OVER)		

(Do Not write in the area below)

*DMB*  
*5/13/11*


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010.

*Daniel E. Nicholas*  
Daniel E. Nicholas *Kyle KLOPPER*  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

MAY 11, 2011

OPERATOR'S NAME:  
RICE DARYL DEAN  
OPERATOR'S NUMBER: 2088  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: UOC  
DEPT/AGENCY: SIT1  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .078  
ALCO S/N: X301572

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.07 in  
ALCO TARGET .078 19:45  
BLANK TEST .000 19:46  
INTERNAL STANDARD VERIFIED 19:46  
ALCO TU 30.07 in .000 19:46  
BLANK TEST .000 19:47  
SUBJECT SAMPLE .000 19:48  
BLANK TEST .000 19:49  
ALCO TU 30.07 in .085 19:51  
BLANK TEST .000 19:52

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

MAY 11, 2011  
TIME 19:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 43c  
BAROMETER: 30.07 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

0127456789:~!@#%&'()\*+,-./:;<=>?@ABCDEF  
GHIJKL\*OPQRSTUVWXYZ~!@#%&'()\*+,-./:;<=>?@ABCDEF  
GHIJKL\*OPQRSTUVWXYZ~!@#%&'()\*+,-./:;<=>?@ABCDEF

*Daryl Rice*  
*05/11/2011*  
*HT PST 250*

*Daryl Rice*  
*05/11/2011*  
*HT PST 250*  
*BMB*  
*5/13/11*

*BMB*  
*5/13/11*