

103082011

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

Supervisor/Operator Performing the Verification Procedure:  
 Name DARYL RICE ID# 2088 Date 08/05/2011

A Agency SITKA POLICE DEPT. Phone # 747-3245  
 Instrument Location 304 LAKE STREET

---

B Alco S/N X172969 Target Value .077 High Pressure 900

Alco Test Values .080 .080  
 1st Alco 2nd Alco

Signature Daryl Rice

(OVER)

*DMB*  
8/8/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2010. *NJB*

*Nikki Roth*  
 (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

AUGUST 05, 2011

OPERATOR'S NAME:  
RICE DARL DEAN  
OPERATOR'S NUMBER: 2088  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :

UOC  
O.L. #: UOC  
DEPT/AGENCY: SIT1  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .077  
ALCO S/N: X172969

### — BREATH ANALYSIS —

.077 ADJUSTED FOR 30.46 in  
ALCO TARGET .078 16:17  
BLANK TEST .000 16:18  
INTERNAL STANDARD VERIFIED 16:18  
ALCO TV 30.46 in .000 16:19  
BLANK TEST .000 16:20  
SUBJECT SAMPLE .000 16:20  
BLANK TEST .000 16:21  
ALCO TV 30.46 in .000 16:21  
BLANK TEST .000 16:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

AUGUST 05, 2011  
TIME 16:23

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.46 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~@