

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

MAY 31 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

|  |                              |   |
|--|------------------------------|---|
| Supervisor/Operator Performing the Verification Procedure: |                              | DataMaster cdm S/N <u>130177</u>                                  |
| Name   | <u>STEVE KEVAN</u>           | ID# <u>6585</u> Date <u>3/11/11</u> <sup>BMB</sup> <u>5/27/11</u> |
| A Agency   | <u>ALASKA STATE TROOPERS</u> | Phone # <u>907-438-2029</u>                                       |
| Instrument Location <u>ST. MARY'S</u>                      |                              |   |
| B Alco S/N   | <u>X172959</u>               | Target Value <u>.081</u> High Pressure <u>600</u>                 |
| Alco Test Values   | <u>.081</u><br>1st Alco      | <u>.081</u><br>2 <sup>nd</sup> Alco                               |
| Signature <u>[Signature]</u>                               |                              |   |

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011 <sup>2011</sup> <sub>MSB</sub>

[Signature] (Notary Seal Stamp)  
KYLE KLOPPER  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177

MAY 27, 2011

OPERATOR'S NAME: KEVIN, STEVE  
OPERATOR'S NUMBER: 6565  
SUBJECT'S LAST NAME: TEST  
SUBJECT'S FIRST NAME/MI: TEST  
Q.L. #: 00  
DEPT/AGENCY: SAKO  
CASE/REPORT: 11-48530  
TEST TYPE: U  
ALCO TARGET VALUE: .081  
ALCO S/N: X172959

— BREATH ANALYSIS —

|                            |          |       |
|----------------------------|----------|-------|
| .081 ADJUSTED FOR 29.88 in |          |       |
| ALCO TARGET                | .080     | 10:42 |
| BLANK TEST                 | .080     | 10:43 |
| INTERNAL STANDARD          | VERIFIED | 10:43 |
| ALCO TV 29.88 in           | .081     | 10:44 |
| BLANK TEST                 | .080     | 10:44 |
| SUBJECT SAMPLE             | .080     | 10:45 |
| BLANK TEST                 | .080     | 10:46 |
| ALCO TV 29.90 in           | .081     | 10:46 |
| BLANK TEST                 | .080     | 10:47 |

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177

MAY 27, 2011  
TIME 10:35

— DIAGNOSTIC CHECK —

|                  |          |
|------------------|----------|
| COMPUTER:        | OKAY     |
| PROGRAM:         | OKAY     |
| SOFTWARE DATE:   | 02/20/01 |
| HEATERS          |          |
| SAMPLE CHAMBER:  | 49c      |
| BREATH TUBE:     | 42c      |
| BAROMETER:       | 29.90 in |
| FLOW DETECTOR:   | OKAY     |
| PUMP             |          |
| HIGH SPEED:      | OKAY     |
| DETECTOR:        | OKAY     |
| FILTERS:         | OKAY     |
| QUARTZ STANDARD: | OKAY     |
| CALIBRATION:     | OKAY     |

PRINTER TEST

!@#\$%^&\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~

