VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

Supervisor/Operator Performing the Verification Procedure:

A

Name Garrett Willis ID# 5560 Date 1/9/09

Agency Alaska State Troopers Phone # 1-800-240-2019

Instrument Location St. Mary's AST

B

Alco S/N X30155 Target Value .079 High Pressure 1000

Alco Test Value Average .082 (1st Alco) .082 (2nd Alco)

Signature Garrett Willis CWB 1/20/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Garrett Willis, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

2/3/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

JAN 13 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

JANUARY 09, 2009 CDB 1/20/09

OPERATOR'S NAME:
WILLIS/GARRETT/R
OPERATOR'S NUMBER: 5560
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: L GRW
DEPT/AGENCY: SMV0 1/9/09
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301550

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.85 in

ALCO TARGET	.078	09:05
BLANK TEST	.000	09:05
INTERNAL STANDARD	VERIFIED	09:06
ALCO TV 29.85 in	.082	09:06
BLANK TEST	.000	09:07
SUBJECT SAMPLE	.000	09:07
BLANK TEST	.000	09:08
ALCO TV 29.85 in	.082	09:09
BLANK TEST	.000	09:09

CDB 1/20/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

JANUARY 09, 2009
TIME 09:01

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY GRW 1/9/09
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.85 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~