

FEB 07 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130177

Name SEAN E ADKINS ID# 5110 Date 2/05/2008

A Agency ALASKA STATE TROOPERS Phone # 1-800-240-2019

Instrument Location ST. MARY'S

B Alco S/N 58428 Target Value .079 High Pressure 800

Alco Test Values

	<u>.081</u>	<u>.081</u>
	1 st Alco	2 nd Alco

Signature [Signature] CSB 2/21/08

Under the Alaska Rules of Evidence, I certify that, (OVER)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Sean Adkins, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature] 02/22/08
 Jeanne Swartz Date
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

FEBRUARY 05, 2008

OPERATOR'S NAME:

ADKINS/S/E

OPERATOR'S NUMBER: 5110

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/NI:

VOC

O.L. #: VOC

DEPT/AGENCY: SMIYB

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: 58428

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.01 in
ALCO TARGET .079 14:42
BLANK TEST .000 14:43
INTERNAL STANDARD (VERIFIED) 14:43
ALCO TV 30.03 in .081 14:43
BLANK TEST .000 14:44
SUBJECT SAMPLE .000 14:45
BLANK TEST .000 14:45
ALCO TV 30.01 in .081 14:46
BLANK TEST .000 14:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

FEBRUARY 05, 2008

TIME 14:39

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.01 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`abcdefg|hijklmno
pqrstuvwxy{ }~