


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 12 2008

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130177 ✓
Name		Sean Adkins	ID# 5110 Date 11/07/2008 ✓
A	Agency	ALASKA STATE TROOPERS	Phone # 438-2029
Instrument Location		St. Mary's ✓	
Alco S/N		X301550 ✓	Target Value .079 ✓ High Pressure 1200 PSI
B	Alco Test Value Average	.082 ✓ 1 st Alco	.081 ✓ 2 nd Alco
Signature			 11/20/08

(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Sean Adkins, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/21/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

NOVEMBER 07, 2000

OPERATOR'S NAME:

ADKINS/S/E

OPERATOR'S NUMBER: 5110

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: VOC

DEPT/AGENCY: SSWD

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: K301550

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.86 in
ALCO TARGET .078 13:41
BLANK TEST .000 13:42
INTERNAL STANDARD VERIFIED 13:42
ALCO TV 29.86 in .082 ✓ 13:43
BLANK TEST .000 13:43
SUBJECT SAMPLE *CUB* .000 13:44
BLANK TEST *11/20/00* .000 ✓ 13:45
ALCO TV 29.88 in .081 13:45
BLANK TEST .000 13:46

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

NOVEMBER 07, 2000 ✓

TIME 13:38

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.86 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
pqrs tuvwxyz{ }~