

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 10 2010

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130177</u>	
Name <u>STEVEN C. KEVAN</u>	ID# <u>6585</u> Date <u>3/7/10</u>
A Agency <u>ALASKA STATE TROOPERS</u> Phone # <u>800-240-2019</u>	
Instrument Location <u>ST. MARY'S</u>	
B Alco S/N <u>X 301550</u> Target Value <u>.099</u> High Pressure <u>500</u>	
Alco Test Value Average <u>.099</u> <u>.099</u>	
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	
(OVER)	

BMB  
3/24/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5<sup>th</sup> day of April, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177

MARCH 07, 2010 ✓

OPERATOR'S NAME:  
KEVIN/STEVE

OPERATOR'S NUMBER: 6585

SUBJECT'S LAST NAME:  
TEST

SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: 00

DEPT/AGENCY: SANY

CASE/REPORT: 00

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301558 ✓

### — BREATH ANALYSIS —

.079 ADJUSTED FOR 29.52 in ✓  
ALCO TARGET .077 10:21  
BLANK TEST .000 10:22  
INTERNAL STANDARD VERIFIED 10:22  
ALCO TU 29.52 in .079 ✓ 10:22  
BLANK TEST .000 10:23  
SUBJECT SAMPLE .000 10:23  
BLANK TEST .000 10:24  
ALCO TU 29.54 in .079 ✓ 10:25  
BLANK TEST .000 10:25

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130177 ✓

MARCH 07, 2010 ✓  
TIME 10:17

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.52 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~