

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 16 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130 178</u>
Name	<u>Trooper James K. Kimura</u>	ID#	<u>5677</u> Date <u>3/13/09</u> ✓
A	Agency	<u>Alaska State Troopers</u>	Phone # <u>225-5118</u>
	Instrument Location	<u>Ketchikan AST Post</u>	✓
B	Alco S/N	<u>X172968</u> ✓	Target Value <u>.083</u> ✓ High Pressure <u>1,200 PSI</u>
	Alco Test Values	<u>.083</u> ✓ 1 <sup>st</sup> Alco	<u>.083</u> ✓ 2 <sup>nd</sup> Alco
	Signature	<u>[Signature]</u> JKMD 3/13/09	<u>CSB</u> <u>3/24/09</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that, <sup>(Do Not write in the area below)</sup>

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Kimura, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/24/09

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 178 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130178

MARCH 13, 2009 ✓ *CSB 3/24/09*

OPERATOR'S NAME:  
KIMURA/JAMES/K  
OPERATOR'S NUMBER: 5677  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
VOC  
O.L. #: 000000  
DEPT/AGENCY: KETA  
CASE/REPORT: 09-000000  
TEST TYPE: V  
ALCO TARGET VALUE: .083 ✓  
ALCO S/N: X172968

### — BREATH ANALYSIS —

.083 ADJUSTED FOR 29.75 in		
ALCO TARGET	.082	21:29
BLANK TEST	.000	21:31
INTERNAL STANDARD	VERIFIED	21:31
ALCO TV 29.75 in	.083 ✓	21:31
BLANK TEST	.000	21:32
SUBJECT SAMPLE <i>CSB 3/24/09</i>	.000	21:32
BLANK TEST	.000 ✓	21:33
ALCO TV 29.75 in	.083	21:34
BLANK TEST	.000	21:34

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130178

MARCH 13, 2009 ✓  
TIME 21:39

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	29.75 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■