

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 15 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130 178</u> ✓
Name <u>James Kimura</u>		ID#	<u>5677</u> Date <u>4/28/09</u> ✓
A	Agency <u>Alaska State Troopers</u>	Phone #	<u>907 225-5118</u>
Instrument Location		<u>Ketchikan Post</u> ✓	
Alco S/N <u>X172968</u> ✓		Target Value	<u>.083</u> ✓ .085
		High Pressure	<u>1,100 PSI</u>
B	Alco Test Values	<u>.084</u> ✓ 1 st Alco	
		<u>.085</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u>		<u>(4/28/09)</u> ✓	
		<u>Bmb</u> <u>5/20/09</u>	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland,
Notary Public, State of Alaska
Commission Expires with Office

