

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 09 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130 178</u>
Name <u>Trooper James K Kimura</u>		ID# <u>5677</u>	Date <u>6/26/09</u>
A	Agency <u>Alaska State Troopers</u>	Phone # <u>225-5118</u>	
Instrument Location <u>Ketchikan AST Post</u>			
Alco S/N <u>X 172968</u>		Target Value <u>.083</u>	High Pressure <u>750 PSI</u>
B	Alco Test Values	<u>.084</u> 1 st Alco	<u>.083</u> 2 nd Alco
Signature <u>[Signature]</u>		<u>CQB</u> <u>7/14/09</u>	

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of 20th, 2009.

[Signature]
Carolyn M. Noiland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 178

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

JUNE 26, 2009

OPERATOR'S NAME:
KIMURA/JAMES/K
OPERATOR'S NUMBER: 5677
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/NI :
UOU

O.L. #: 0000000
DEPT/AGENCY: KETO
CASE REPORT: 09-00000
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: X172968

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.09 in
ALCO TARGET .083 20:05
BLANK TEST .000 20:06
INTERNAL STANDARD VERIFIED 20:06
ALCO TV 30.09 in .084 20:07
BLANK TEST .000 20:07
SUBJECT SAMPLE .000 20:08
BLANK TEST .000 20:08
ALCO TV 30.09 in .083 20:09
BLANK TEST .000 20:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

JUNE 26, 2009
TIME 20:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.09 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL MNOPQRSTU VWXYZ[\]^_`abdefghijklmno
pqrstuvwxy z{|}~