

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

DEC 18 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130178 BMB</u> ✓ 103-178
Name	<u>James K. Kimura</u>	ID# <u>5677</u> Date <u>12/13/09</u>
A Agency	<u>Alaska State Troopers</u>	Phone # <u>225-5118</u>
Instrument Location	<u>Ketchikan AST Post</u> ✓	
B Alco S/N	<u>X124662</u>	Target Value <u>.080</u> High Pressure <u>300 PSI</u>
Alco Test Values	<u>.084</u> ✓ 1 st Alco	<u>.085</u> ✓ 2 nd Alco
Signature	<u>[Signature]</u> ✓	

(OVER) BMB
1/5/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of Jan 2010

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N ~~103 178~~ ¹³⁰¹⁷⁸ BMB

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

[]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

DECEMBER 13, 2009 ✓ BMB 11510

OPERATOR'S NAME:
KIMURA/JAMES/K
OPERATOR'S NUMBER: 5677
SUBJECT'S LAST NAME:
UC
SUBJECT'S FIRST NAME/MI :
UC
O.L. #: 000000
DEPT/AGENCY: KETO
CASE/REPORT: 09-000000
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X124662 ✓

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.14 in ✓
ALCO TARGET .000 21:39
BLANK TEST .000 21:41
INTERNAL STANDARD VERIFIED 21:41
ALCO TV 30.14 in .004 ✓ 21:41
BLANK TEST .000 21:42
SUBJECT SAMPLE BMB .000 21:42
BLANK TEST .000 21:43
ALCO TV 30.14 in .005 21:44
BLANK TEST .000 21:45

[]

DIAGNOSTIC CHECK

[]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178 ✓

DECEMBER 13, 2009 ✓
TIME 21:47

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.14 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

[]