

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | |
|--------------------------------------------------------------------------|-------------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | |
| Name <u>DAN STURKIE</u> | ID# <u>3559</u> Date <u>02-7-11</u> |
| A Agency <u>Anchorage Police</u> | Phone # <u>786-8900</u> |
| Instrument Location <u>Eagle River Substation</u> | |
| Alco S/N <u>75260</u> Target Value <u>0.081</u> High Pressure <u>500</u> | |
| B Alco Test Values | <u>0.083</u> ✓ 1st Alco |
| | <u>0.083</u> ✓ 2nd Alco |
| Signature <u>D.R. Sturkie</u> | |

(OVER)

(Do Not write in the area below)

DMB
3/6/11


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of MAR, 2011.

[Signature]
(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130178

(CONTINUED FROM FRONT PAGE)

82-7-11

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

FEBRUARY 07, 2011

OPERATOR'S NAME:

STUEKIW

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

V

O.L. #: VOC

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 75260

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.38 in

ALCO TARGET .082

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 30.38 in .083

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 30.40 in .083

BLANK TEST .000

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

FEBRUARY 07, 2011

TIME 23:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.38 in

FLOW DETECTOR: OKAY

PUMP

23:15 HIGH SPEED: OKAY

23:15 DETECTOR: OKAY

23:16 DETECTOR: OKAY

23:16 FILTERS: OKAY

23:17 FILTERS: OKAY

23:18 QUARTZ STANDARD: OKAY

23:18 QUARTZ STANDARD: OKAY

23:19 CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDE
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklm
pqrstuvwxyz{|}~△