

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>132178</u> ✓
Name	<u>C. RIVALA</u>	ID# <u>3791</u> ✓ Date <u>3/24/11</u> ✓
A Agency	<u>APD</u>	Phone # _____
Instrument Location <u>WOLLAN SUB, EAGLE RIVER</u>		
B Alco S/N	<u>75260</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>500</u>
Alco Test Values	<u>081</u> ✓ 1 st Alco	<u>082</u> ✓ 2 nd Alco
Signature	<u>[Signature]</u>	<u>Rivala/1349</u>

(OVER)

BMB
3/31/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 1ST day of APRIL, 2011.

[Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130178

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

MARCH 24, 2011

OPERATOR'S NAME: RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
SIXTY DAY
O.L. #: A
DEPT/AGENCY: ANCL
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: 75260

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.55 in
ALCO TARGET .080 18:33
BLANK TEST .000 18:33
INTERNAL STANDARD VERIFIED 18:33
ALCO TV 29.55 in .081 18:34
BLANK TEST .000 18:34
SUBJECT SAMPLE .000 18:35
BLANK TEST .000 18:36
ALCO TV 29.57 in .082 18:36
BLANK TEST .000 18:37

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

MARCH 24, 2011
TIME 18:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~Δ