VERIFICATION OF CALIBRATION REPORT FEB 25 2009

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

Supervisor/Operator Performing the Verification Procedure:

A

Name John T. Ostoj ID# 4823 Date 2/20/09

Agency AST Phone #

Instrument Location Soldotna AST

B

Alco S/N CCB 3/2/09 X58442 Target Value .080 High Pressure 1200

Alco Test Value Average .084 (1st Alco) .084 (2nd Alco)

Signature [signature] CCB 3/10/09

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 3/11/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

FEBRUARY 20, 2009

OPERATOR'S NAME: CCB 3/10/09
OSTOJ/JOHN/T
OPERATOR'S NUMBER: 4823
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC TEST
DEPT/AGENCY: SLD0
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: 458442
CCB 3/2/09

— BREATH ANALYSIS —

.080 ADJUSTED FOR 30.71 in

ALCO TARGET	.082	11:56
BLANK TEST	.000	11:57
INTERNAL STANDARD	VERIFIED	11:57
ALCO TV 30.71 in	.084	11:57
BLANK TEST	.000	11:58
SUBJECT SAMPLE CCB 3/10/09	.000	11:59
BLANK TEST	.000	12:00
ALCO TV 30.71 in	.084	12:00
BLANK TEST	.000	12:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

FEBRUARY 20, 2009
TIME 11:54

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	30.71 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~