

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130179</u> ✓
Name		<u>John T. Ostoj</u> ✓	ID# <u>4823</u> Date <u>4-29-08</u> ✓
A	Agency	<u>Soldotna AST</u> ✓	Phone # <u>262-4453</u>
Instrument Location		<u>Soldotna AST Post</u>	
Alco S/N		<u>X124658</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>200</u>
B	Alco Test Value Average	<u>.079</u> ✓ 1 st Alco	<u>.079</u> ✓ 2 nd Alco
Signature		<u><i>John T. Ostoj</i></u>	<u>COB</u> <u>5/5/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that (I am the operator in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

5/5/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm-breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

APRIL 29, 2008 *08/26/08*

OPERATOR'S NAME:
OSTOJ/JOHN/T
OPERATOR'S NUMBER: 4823
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
TEST

O.L. #: UOC TEST
DEPT/AGENCY: SLDG
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: X124658

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.06 in
ALCO TARGET .081 08:32
BLANK TEST .000 08:33
INTERNAL STANDARD VERIFIED 08:33
ALCO TV 30.06 in .079 08:34
BLANK TEST .000 08:35
SUBJECT SAMPLE .000 *08/26/08* 08:35
BLANK TEST .000 08:36
ALCO TV 30.06 in .079 08:36
BLANK TEST .000 08:37

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

APRIL 29, 2008
TIME 08:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.06 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`abcde fgh i jk lmn o
pqrstu vwxyz{|}~