

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130179 ✓

Name V. Eugene Fowler ID# 4386 Date 6-9-10

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna AST Post

B Alco S/N X173010 Target Value .085 High Pressure 1200

Alco Test Value Average .084 .084
1st Alco 2nd Alco

Signature V E Fowler CCB
6/21/10

(OVER)

Mailed
6-9-10
VEF

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of July 2010.
Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179 ✓

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

JUNE 09, 2010 ✓
COB 6/2/10

OPERATOR'S NAME: FOWLER/A/E
OPERATOR'S NUMBER: 4386
SUBJECT'S LAST NAME: UOC
SUBJECT'S FIRST NAME/MI: UOC
O.L. #: UOC
DEPT/AGENCY: SLT5
CASE/REPORT: UOC
TEST TYPE: U ✓
ALCO TARGET VALUE: .085 ✓
ALCO S/N: X173010 ✓

— BREATH ANALYSIS —

.085 ADJUSTED FOR 30.19 in		
ALCO TARGET	.085	14:49
BLANK TEST	.000	14:50
INTERNAL STANDARD	VERIFIED	14:50
ALCO TV 30.19 in	.004 ✓	14:50
BLANK TEST	.000	14:51
SUBJECT SAMPLE	.000 ✓	14:52
BLANK TEST	.000 ✓	14:53
ALCO TV 30.19 in	.004	14:53
BLANK TEST	.000	14:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179 ✓

JUNE 09, 2010 ✓
TIME 14:42

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.19 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`abcde fghijklmno
pqrstuvwxyz{|}~||