

VERIFICATION OF CALIBRATION REPORT JUL 08 2010
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179[✓]

Supervisor/Operator Performing the Verification Procedure:

Name V. Eugene Fowler ID# 4386[✓] Date 7-5-10

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna AST Post

B Alco S/N 78259[✓] Target Value .083[✓] High Pressure 1400

Alco Test Value Average .083[✓] .083[✓]
 1st Alco 2nd Alco

Signature *V. Eugene Fowler*

(OVER)

BMB
7/15/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 27th day of July, 2010.

Carolyn M. Noland (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

JULY 05, 2010

OPERATOR'S NAME: *BMB*
FOWLER/U/E
OPERATOR'S NUMBER: 4396
SUBJECT'S LAST NAME: *7/15/10*

UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC
DEPT/AGENCY: SLT5
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: 78259

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .083 ADJUSTED FOR 30.03 in | | |
| ALCO TARGET | .083 | 19:00 |
| BLANK TEST | .000 | 19:02 |
| INTERNAL STANDARD | VERIFIED | 19:02 |
| ALCO TV 30.03 in | .083 | 19:02 |
| BLANK TEST | .000 | 19:03 |
| SUBJECT SAMPLE | .000 | 19:04 |
| BLANK TEST | .000 | 19:04 |
| ALCO TV 30.04 in | .083 | 19:05 |
| BLANK TEST | .000 | 19:06 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

JULY 05, 2010
TIME 19:09

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 41c |
| BAROMETER: | 30.04 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL^NOPQRSTUVWXYZ[\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~#