


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130179 ✓
Name		John T. Ostoj	ID# 4823 Date 7-26-08 ✓
A	Agency	AST Soldotna	Phone # 262-4453
Instrument Location		Soldotna ASF Post ✓	
Alco S/N		X172946 ✓	Target Value .076 ✓ High Pressure 1000
B	Alco Test Value Average	.076 ✓ 1 st Alco	.077 ✓ 2 nd Alco
Signature			COB 9/11/08
(OVER)			



Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

9/11/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number 130179 ✓ Model DataMaster
Location SLD0 ✓ AST/Soldotna ✓

CSB
9/11/08

Diagnostic Test Record

Serial Number 130179 ✓ Model DataMaster
Location SLD0 ✓ AST/Soldotna

Subject Information

Last Name Tst O.L. # J
First Name Test

Diagnostic Test Results

Test Date 07/26/2008 ✓
Test Time 14:01:43
Computer Okay
Program Okay
Software Date 02/20/2001
Sample Chamber (C) 49
Breath Tube (C) 40
Barometer (in) 29.93
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay

Operator Information

Operator Name Dstoj ✓ Case D
Operator # 4823 Department SLD0

Test Results

Alco Target .076 ✓ Alco Serial X172946 ✓
Nominal 0.076 Barometer 29.93 Test Type T
Dry Gas 1 0.076 ✓ Time: 14:05:10 Test Date 07/26/2008 ✓
Subject Alcohol 0.000 Time: 14:06:34 Test Time 14:03:47
Dry Gas 2 0.077 ✓ Time: 14:07:41 Report Y