VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

Supervisor/Operator Performing the Verification Procedure:

A

Name Ostoj, John T. ID# 4823 Date 9-27-08

Agency Soldotna AST Phone # 262-4453

Instrument Location Soldotna AST Post

B

Alco S/N X301556 Target Value .081 High Pressure 1100

Alco Test Value Average .080 (1st Alco) .079 (2nd Alco)

Signature [signature] 11/20/08

(OVER)

(Do Not Write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 11/21/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

OCT 06 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

SEPTEMBER 27, 2008

OPERATOR'S NAME:
OSTOJ/JOHN/T
OPERATOR'S NUMBER: 4823
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: TEST
DEPT/AGENCY: SL00
CASE/REPORT: TEST
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X301556

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.93 in

ALCO TARGET	.081	07:17
BLANK TEST	.000	07:18
INTERNAL STANDARD	VERIFIED	07:18
ALCO TV 29.93 in	.080	07:19
BLANK TEST	.000	07:20
SUBJECT SAMPLE	.000	07:20
BLANK TEST	.000	07:21
ALCO TV 29.93 in	.079	07:21
BLANK TEST	.000	07:22

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

SEPTEMBER 27, 2008
TIME 07:24

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.94 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~