

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 06 2011

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130179

Name V. Eugene Fowler ID# 4386 Date 3/17/11

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna AST

---

B Alco S/N 718493 Target Value .080 High Pressure 800

Alco Test Values .079 .077

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

[Handwritten initials]


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011. NJB

[Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas Kyle Klopfer  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130179

MARCH 17, 2011

OPERATOR'S NAME:

FOWLER/V/E

OPERATOR'S NUMBER: 4386

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: SLT5

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .080

ALCO S/N: 718493

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.16 in  
ALCO TARGET .080 15:20  
BLANK TEST .000 15:21  
INTERNAL STANDARD VERIFIED 15:21  
ALCO TV 30.16 in .079 15:21  
BLANK TEST .000 15:22  
SUBJECT SAMPLE .000 15:23  
BLANK TEST .000 15:23  
ALCO TV 30.16 in .077 15:24  
BLANK TEST .000 15:25

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130179

MARCH 17, 2011

TIME 15:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.16 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ\]^\_`abcde fghi jklmno  
pqrstuvwxy z{|}~■