

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130179</u> ✓
Name		<u>John T. Ostoj</u>	ID# <u>4823</u> Date <u>12/22/08</u> ✓
A	Agency	<u>Soldotna AST</u>	Phone # _____
Instrument Location		<u>Soldotna AST</u> ✓	
Alco S/N		<u>X301755</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>950</u>
B	Alco Test Value Average	<u>.080</u> ✓ 1 st Alco	<u>.079</u> ✓ 2 nd Alco
Signature		<u>John T. Ostoj</u> ✓	<u>CSB</u> <u>12/30/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/31/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

DECEMBER 22, 2008

OPERATOR'S NAME: *CBS 12/20/08*
OSTOJ/JOHN/T
OPERATOR'S NUMBER: 4923
SUBJECT'S LAST NAME: TEST
SUBJECT'S FIRST NAME/MI: TEST
O.L. #: TEST
DEPT/AGENCY: SLDG
CASE/REPORT: TEST
TEST TYPE: U
ALCO TARGET VALUE: .090 ✓
ALCO S/N: 1301755

--- BREATH ANALYSIS ---

.090 ADJUSTED FOR 30.07 in		
ALCO TARGET	.090	09:44
BLANK TEST	.000	09:46
INTERNAL STANDARD	VERIFIED	09:46
ALCO TO 30.07 in	.090 ✓	09:46
BLANK TEST	.000	09:47
SUBJECT SAMPLE <i>CBS 12/20/08</i>	.090	09:47
BLANK TEST	.000	09:48
ALCO TO 30.07 in	.079 ✓	09:49
BLANK TEST	.000	09:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

DECEMBER 22, 2008 ✓
TIME 09:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 49c

BAROMETER: 30.07 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARDS: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/17/07 10:11:21 AM / 01234567891011 = 1234567891011
ALCO DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY