

*of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Datamaster cdm S/N 130180

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

APRIL 02, 2009

OPERATOR'S NAME: *COB*
ANDERSON AM *Slain*

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:
SIXTY DAY VERIFICATION

SUBJECT'S FIRST NAME/MI :
N/A

O.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X301530

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.20 in		
ALCO TARGET	.078	12:33
BLANK TEST	.000	12:34
INTERNAL STANDARD	VERIFIED	12:35
ALCO TU 30.20 in	.076 ✓	12:35
BLANK TEST	.000	12:36
SUBJECT SAMPLE <i>COB</i>	.000	12:36
BLANK TEST	.000	12:37
ALCO TU 30.20 in	.076 ✓	12:37
BLANK TEST	.000	12:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

APRIL 02, 2009 ✓

TIME 12:38

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.20 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~

VERIFICATION OF CALIBRATION REPORT 1

of Datamaster cdm
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

A

Supervisor/Operator Performing the Verification Procedure: Datamaster cdm S/N 130180

Name AW Anderson ID# 2467 Date 4/02/09

Agency Seldovia Police Dept. Phone # 234-7640

Instrument Location Seldovia Police Dept.

B

Alco S/N X301530 Target Value .018 High Pressure 100 PSI

Alco Test Value Average .076 .076
1st Alco 2nd Alco

Signature AW Anderson COS
5/13/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of May, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

