

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130180 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

JULY 07, 2008 *COB 7/10/08*

OPERATOR'S NAME:
ANDERSON, AM
OPERATOR'S NUMBER: 2467
SUBJECT'S LAST NAME:
SIXTY DAY VERIFICATION
SUBJECT'S FIRST NAME/MI :
N/A

D.L. #: 1234567890
DEPT/AGENCY: SOVI
CASE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .078 ✓
ALCO S/N: X301530 ✓

--- BREATH ANALYSIS ---

| | | |
|----------------------------|---------------------|-------|
| .078 ADJUSTED FOR 30.12 in | | |
| ALCO TARGET | .078 | 11:56 |
| BLANK TEST | .000 | 11:57 |
| INTERNAL STANDARD | VERIFIED | 11:57 |
| ALCO TV 30.11 in | .075 ✓ | 11:57 |
| BLANK TEST | .000 <i>COB</i> | 11:58 |
| SUBJECT SAMPLE | .000 <i>COB</i> | 11:59 |
| BLANK TEST | .000 <i>7/10/08</i> | 12:00 |
| ALCO TV 30.12 in | .075 | 12:01 |
| BLANK TEST | .000 | 12:01 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

JULY 07, 2008 ✓
TIME 12:02

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 41c |
| BAROMETER: | 30.12 in |
| FLOW DETECTOR: | OKAY |
| FLAMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk l m n o
pqrstuvwxyz{|}~#